


# IDKmonitor<sup>®</sup> Infliximab drug level ELISA

*For the determination of free infliximab  
concentration (e.g. REMICADE<sup>®</sup>, Remsima<sup>®</sup>, Inflectra<sup>®</sup>)  
in EDTA plasma and serum*

Valid from 2022-08-09

REF KR9655

$\Sigma$  96



RUO

REF KR9655.20

$\Sigma$  20 x 96



**Immundiagnostik AG**, Stubenwald-Allee 8a, 64625 Bensheim, Germany

Tel.: +49 6251 70190-0

Fax: + 49 6251 70190-363

e.mail: [info@immundiagnostik.com](mailto:info@immundiagnostik.com)

[www.immundiagnostik.com](http://www.immundiagnostik.com)

# Table of Contents

<b>1. INTENDED USE</b>	<b>2</b>
<b>2. INTRODUCTION</b>	<b>2</b>
<b>3. MATERIAL SUPPLIED</b>	<b>2</b>
<b>4. MATERIAL REQUIRED BUT NOT SUPPLIED</b>	<b>3</b>
<b>5. STORAGE AND PREPARATION OF REAGENTS</b>	<b>3</b>
<b>6. STORAGE AND PREPARATION OF SAMPLES</b>	<b>4</b>
<b>7. ASSAY PROCEDURE</b>	<b>4</b>
<i>Principle of the test</i>	4
<i>Test procedure</i>	5
<b>8. RESULTS</b>	<b>6</b>
<b>9. LIMITATIONS</b>	<b>7</b>
<b>10. QUALITY CONTROL</b>	<b>7</b>
<i>Reference range</i>	7
<b>11. PERFORMANCE CHARACTERISTICS</b>	<b>7</b>
<i>Accuracy – Precision</i>	7
<i>Analytical sensitivity</i>	8
<b>12. PRECAUTIONS</b>	<b>8</b>
<b>13. TECHNICAL HINTS</b>	<b>9</b>
<b>14. GENERAL NOTES ON THE TEST AND TEST PROCEDURE</b>	<b>9</b>
<b>15. REFERENCES</b>	<b>10</b>

## 1. INTENDED USE

The Immundiagnostik AG assay is an enzyme immunoassay intended for the quantitative determination of free therapeutic TNF $\alpha$  antibodies infliximab (e.g. REMICADE®, Remsima®, Inflectra®) in EDTA plasma and serum. For research use only. Not for use in diagnostic procedures.

## 2. INTRODUCTION

Tumor necrosis factor alpha (TNF $\alpha$ ) belongs to the proinflammatory cytokines, which promote and sustain inflammatory reactions. It is produced by macrophages and T cells and plays a central role in both acute and chronic inflammations.

The IDKmonitor® infliximab drug level ELISA for the determination of the drug level of infliximab measures quantitatively free infliximab in EDTA plasma and serum.

## 3. MATERIAL SUPPLIED

Cat. No.	Label	Kit Components	Quantity for cat. no.	
			KR9655	KR9655.20
KR9655	PLATE	Microtiter plate, pre-coated	12 x 8 wells	20 x 12 x 8 wells
KR0001.C.100	WASH-BUF	Wash buffer concentrate, 10x	2 x 100 ml	20 x 100 ml
KR9655	CONJ	Conjugate concentrate, peroxidase-labelled	1 x 200 $\mu$ l	20 x 200 $\mu$ l
KR9655	STD	Calibrators, lyophilised (see specification for concentrations)	4 x 6 vials	40 x 6 vials
KR9655	CTRL 1	Control, lyophilised (see specification for range)	4 x 1 vial	40 x 1 vial
KR9655	CTRL 2	Control, lyophilised (see specification for range)	4 x 1 vial	40 x 1 vial
KR0004.100	SAMPLE-BUF	Sample dilution buffer, ready-to-use	2 x 100 ml	40 x 100 ml
KR0002.15	SUB	Substrate (tetramethylbenzidin), ready-to-use	1 x 15 ml	20 x 15 ml

Cat. No.	Label	Kit Components	Quantity for cat. no.	
			KR9655	KR9655.20
KR0003.15	STOP	Stop solution, ready-to-use	1 x 15 ml	20 x 15 ml

For reorders of single components, use the catalogue number followed by the label as product number.

#### 4. MATERIAL REQUIRED BUT NOT SUPPLIED

- Ultrapure water\*
- Calibrated precision pipettors and 10–1000 µl single-use tips
- Absorbent paper
- Foil to cover the microtiter plate
- Horizontal microtiter plate shaker
- Multi-channel pipets or repeater pipets
- Centrifuge
- Vortex
- Standard single-use laboratory glass or plastic vials, cups, etc.
- Microtiter plate reader (required filters see chapter 7)

\* Immundiagnostik AG recommends the use of ultrapure water (water type 1; ISO 3696), which is free of undissolved and colloidal ions and organic molecules (free of particles >0.2 µm) with an electrical conductivity of 0.055 µS/cm at 25 °C (≥ 18.2 MΩ cm).

#### 5. STORAGE AND PREPARATION OF REAGENTS

- To run the assay more than once, ensure that reagents are stored at the conditions stated on the label. **Prepare only the appropriate amount necessary for each run.** The kit can be used up to 4 times within the expiry date stated on the label.
- Reagents with a volume less than **100 µl** should be centrifuged before use to avoid loss of volume.
- **Preparation of the wash buffer:** The **wash buffer concentrate (WASHBUF)** has to be diluted with ultrapure water **1:10** before use (100 ml WASHBUF + 900 ml ultrapure water), mix well. Crystals could occur due to high salt concentration in the concentrate. Before dilution, the crystals have to be redissolved at room temperature or in a water bath at 37 °C. The **WASHBUF** is stable at **2–8 °C** until the expiry date stated on the label. **Wash buffer** (1:10 diluted WASHBUF) can be stored in a closed flask at **2–8 °C for 1 month**.

- The **lyophilised standards (STD)** and **controls (CTRL)** can be used until the expiry date stated on the label when stored at **2–8 °C**. Before use, the STD and CTRL have to be reconstituted with **500 µl of ultrapure water** and mixed by gentle inversion to ensure complete reconstitution. Allow the vial content to dissolve for 10 minutes and then mix thoroughly. **Standards and controls** (reconstituted STD and CTRL) **can be stored at -20 °C for 3 months. Avoid repeated thawing and freezing.**
- **Preparation of the conjugate:** Before use, the **conjugate concentrate (CONJ)** has to be diluted **1:101** in **wash buffer** (100 µl CONJ + 10 ml wash buffer). The CONJ is stable at **2–8 °C** until the expiry date stated on the label. **Conjugate (1:101 diluted CONJ) is not stable and cannot be stored.**
- All other test reagents are ready-to-use. Test reagents are stable until the expiry date (see label) when stored at **2–8 °C**.

## 6. STORAGE AND PREPARATION OF SAMPLES

### Sample storage

Freshly collected EDTA plasma or serum can be stored for 7 days at room temperature (15–30 °C) [14] or at 2–8 °C. Long term storage is recommended at -20 °C.

Diluted EDTA plasma or serum samples can be stored for 7 days at room temperature, 15 days at 2–8 °C and at least for 7 weeks at -20 °C. More than 3 freeze-thaw cycles are to be avoided.

### EDTA plasma and serum

EDTA plasma or serum samples must be diluted **1:200** before performing the assay, e.g.

**10 µl** sample + **1990 µl** dilution buffer (SAMPLEBUF), mix well.

For testing in duplicates, pipet **2 x 100 µl** per well of each prepared sample.

## 7. ASSAY PROCEDURE

### *Principle of the test*

This ELISA is designed to determine the quantity of free infliximab (therapeutic antibody against TNFα) in EDTA plasma or serum samples. In a first incubation step, the free infliximab from the sample is bound to the specific monoclonal anti-infliximab antibody coated on the plate. To remove all unbound substances, a washing step is carried out. In a further incubation step, peroxidase-labelled antibody is added. Te-

tramethylbenzidine (TMB) is used as a substrate for peroxidase. Finally, an acidic stop solution is added to terminate the reaction. The colour changes from blue to yellow. The intensity of the yellow colour is directly proportional to the concentration of free infliximab in the sample. A dose response curve of the absorbance unit (optical density, OD) vs. concentration is generated, using the values obtained from standard. The concentrations of free infliximab in the samples are determined directly from this curve.

### Test procedure

Bring all **reagents and samples to room temperature** (15–30°C) and mix well.

Mark the positions of standards/controls/samples on a protocol sheet.

Take as many microtiter strips as needed from the kit. Store unused strips together with the desiccant bag in the closed aluminium packaging at 2–8°C. Strips are stable until expiry date stated on the label.

For automated ELISA processors, the given protocol may need to be adjusted according to the specific features of the respective automated platform. For further details please contact your supplier or Immundiagnostik AG.

We recommend to carry out the tests in duplicate.

1.	Add each <b>100 µl standards/controls/diluted samples</b> into the respective wells.
2:	Cover the strips and incubate for <b>1 hour</b> at room temperature (15–30°C) on a <b>horizontal shaker*</b> .
3.	Discard the content of each well and wash <b>5 times</b> with <b>250 µl wash buffer</b> . After the final washing step, remove residual wash buffer by firmly tapping the plate on absorbent paper.
4.	Add <b>100 µl conjugate</b> (diluted CONJ) into each well.
5.	Cover the strips and incubate for <b>1 hour</b> at room temperature (15–30°C) on a <b>horizontal shaker*</b> .
6.	Discard the content of each well and wash <b>5 times</b> with <b>250 µl wash buffer</b> . After the final washing step, remove residual wash buffer by firmly tapping the plate on absorbent paper.
7.	Add <b>100 µl substrate</b> (SUB) into each well.
8.	Incubate for <b>10–20 min**</b> at room temperature (15–30°C) <b>in the dark</b> .
9.	Add <b>100 µl stop solution</b> (STOP) into each well, mix.

10.	Determine <b>absorption immediately</b> with an ELISA reader at <b>450 nm</b> against 620 nm (or 690 nm) as a reference. If no reference wavelength is available, read only at 450 nm. If the extinction of the highest standard exceeds the range of the photometer, absorption must be measured immediately at <b>405 nm</b> against 620 nm as a reference.
-----	---

\* We recommend shaking the strips at 550 rpm with an orbit of 2 mm.

\*\* The intensity of the colour change is temperature sensitive. We recommend observing the colour change and stopping the reaction upon good differentiation.

## 8. RESULTS

The following algorithms can be used alternatively to calculate the results. We recommend using the “4 parameter algorithm”.

### 1. 4 parameter algorithm

It is recommended to use a linear ordinate for the optical density and a logarithmic abscissa for the concentration. When using a logarithmic abscissa, the zero standard must be specified with a value less than 1 (e.g. 0.001).

### 2. Point-to-point calculation

We recommend a linear ordinate for the optical density and a linear abscissa for the concentration.

### 3. Spline algorithm

We recommend a linear ordinate for the optical density and a linear abscissa for the concentration.

The plausibility of the duplicate values should be examined before the automatic evaluation of the results. If this option is not available with the programme used, the duplicate values should be evaluated manually.

### EDTA-plasma and serum samples

The obtained results have to be multiplied by the **dilution factor of 200** to get the actual concentrations.

In case **another dilution factor** has been used, multiply the obtained result by the dilution factor used.

## 9. LIMITATIONS

Samples with concentrations above the measurement range (see definition below) can be further diluted and re-assayed. Please consider this higher dilution when calculating the results.

Samples with concentrations lower than the measurement range (see definition below) cannot be clearly quantified.

The upper limit of the measurement range can be calculated as:

*highest concentration of the standard curve × sample dilution factor to be used*

The lower limit of the measurement range can be calculated as:

*LoB × sample dilution factor to be used*

LoB see chapter "Performance Characteristics".

## 10. QUALITY CONTROL

Immundiagnostik AG recommends the use of external controls for internal quality control, if possible.

Control samples should be analysed with each run. Results, generated from the analysis of control samples, should be evaluated for acceptability using appropriate statistical methods. The results for the samples may not be valid if within the same assay one or more values of the quality control sample are outside the acceptable limits.

### *Reference range*

We recommend each laboratory to establish its own reference range.

## 11. PERFORMANCE CHARACTERISTICS

### *Accuracy – Precision*

#### **Repeatability (Intra-Assay); n = 24**

The repeatability was assessed with 3 serum samples under constant parameters (same operator, instrument, day and kit lot).

Sample	Mean value [µg/ml]	CV [%]
1	4.12	1.9
2	7.88	2.5
3	23.74	7.5



### Reproducibility (Inter-Assay)

The reproducibility was assessed with 6 serum samples under varying parameters (different operators, instruments, days and kit lots).

Sample	Mean value [ $\mu\text{g/ml}$ ]	CV [%]
1 (n = 28)	20.66	12.9
2 (n = 28)	7.79	5.4
3 (n = 22)	17.30	12.0
4 (n = 22)	9.62	7.1
5 (n = 22)	5.13	7.6
6 (n = 22)	2.76	7.6

### Analytical sensitivity

The following values have been estimated based on the concentrations of the standard without considering possibly used sample dilution factors.

Limit of blank, LoB	1.998 ng/ml
Limit of detection, LoD	2.682 ng/ml
Limit of quantitation, LoQ	2.682 ng/ml

The evaluation was performed according to the CLSI guideline EP-17-A2. The specified accuracy goal for the LoQ was 20 % CV.

## 12. PRECAUTIONS

- All reagents in the kit package are for research use only.
- Human materials used in kit components were tested and found to be negative for HIV, Hepatitis B and Hepatitis C. However, for safety reasons, all kit components should be treated as potentially infectious.
- Kit reagents contain sodium azide or ProClin as bactericides. Sodium azide or ProClin are hazardous to health and the environment. Substrates for enzymatic colour reactions may also cause skin and/or respiratory irritation. Any contact with the substances must be avoided. Further safety information can be found in the safety data sheet, which is available from Immundiagnostik AG on request.

- The 10x Wash buffer concentrate (WASHBUF) contains surfactants which may cause severe eye irritation in case of eye contact.

**Warning:** Causes serious eye irritation. **IF IN EYES:** Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing. If eye irritation persists: get medical Advice/attention.

- The stop solution consists of diluted sulphuric acid, a strong acid. Although diluted, it still should be handled with care. It can cause burns and should be handled with gloves, eye protection, and appropriate protective clothing. Any spill should be wiped up immediately with copious quantities of water. Do not breath vapour and avoid inhalation.

### 13. TECHNICAL HINTS

- Do not interchange different lot numbers of any kit component within the same assay. Furthermore we recommend not assembling wells of different microtiter plates for analysis, even if they are of the same batch.
- Control samples should be analysed with each run.
- Reagents should not be used beyond the expiration date stated on the kit label.
- Substrate solution should remain colourless until use.
- To ensure accurate results, proper adhesion of plate sealers during incubation steps is necessary.
- Avoid foaming when mixing reagents.
- Do not mix plugs and caps from different reagents.
- The assay should always be performed according to the enclosed manual.

### 14. GENERAL NOTES ON THE TEST AND TEST PROCEDURE

- The guidelines for laboratories should be followed.
- *IDKmonitor*® is a trademark of Immundiagnostik AG.
- Incubation time, incubation temperature and pipetting volumes of the components are defined by the producer. Any variation of the test procedure, which is not coordinated with the producer, may influence the results of the test. Immundiagnostik AG can therefore not be held responsible for any damage resulting from incorrect use.













- Warranty claims and complaints regarding deficiencies must be logged within 14 days after receipt of the product. The product should be sent to Immunodiagnostik AG along with a written complaint.

## 15. REFERENCES

1. Afif W, Loftus E V, Faubion WA, Kane S V, Bruining DH, Hanson KA, et al. Clinical utility of measuring infliximab and human anti-chimeric antibody concentrations in patients with inflammatory bowel disease. *The American journal of gastroenterology*. 2010/02/11. 2010 May;**105**(5):1133–9.
2. Beglinger C, Binek J, Braegger C, Michetti P, Rogler G, Sauter B, et al. Infliximab-Monotherapie versus Kombinationstherapie mit Immunmodulatoren. *The medical journal*. 2008;**1**:32–4.
3. Bender NK, Heilig CE, Dröll B, Wohlgemuth J, Armbruster F-P, Heilig B. Immunogenicity, efficacy and adverse events of adalimumab in RA patients. *Rheumatology international*. 2006/09/29. 2007 Jan **11**;27(3):269–74.
4. Bendtzen K, Geborek P, Svenson M, Larsson L, Kapetanovic MC, Saxne T. Individualized monitoring of drug bioavailability and immunogenicity in rheumatoid arthritis patients treated with the tumor necrosis factor alpha inhibitor infliximab. *Arthritis and rheumatism*. 2006 Dec;**54**(12):3782–9.
5. Bradley JR. TNF-mediated inflammatory disease. *The Journal of pathology*. 2008 Jan;**214**(2):149–60.
6. St Clair EW, Wagner CL, Fasanmade A a, Wang B, Schaible T, Kavanaugh A, et al. The relationship of serum infliximab concentrations to clinical improvement in rheumatoid arthritis: results from ATTRACT, a multicenter, randomized, double-blind, placebo-controlled trial. *Arthritis and rheumatism*. 2002 Jun;**46**(6):1451–9.
7. Chang JT, Lichtenstein GR. Drug insight: antagonists of tumor-necrosis factor-alpha in the treatment of inflammatory bowel disease. *Nature clinical practice Gastroenterology & hepatology*. 2006 Apr;**3**(4):220–8.
8. Colombel J-F, Loftus E V, Tremaine WJ, Egan LJ, Harmsen WS, Schleck CD, et al. The safety profile of infliximab in patients with Crohn's disease: the Mayo clinic experience in 500 patients. *Gastroenterology*. 2004 Jan;**126**(1):19–31.
9. Cominelli F. Cytokine-based therapies for Crohn's disease--new paradigms. *The New England journal of medicine*. 2004 Nov 11;**351**(20):2045–8.
10. Cornillie F, Shealy D, D'Haens G, Geboes K, Van Assche G, Ceuppens J, et al. Infliximab induces potent anti-inflammatory and local immunomodulatory activity

- but no systemic immune suppression in patients with Crohn's disease. *Alimentary pharmacology & therapeutics*. 2001 Apr;**15**(4):463–73.
11. Maser E a, Vilella R, Silverberg MS, Greenberg GR. Association of trough serum infliximab to clinical outcome after scheduled maintenance treatment for Crohn's disease. *Clinical gastroenterology and hepatology : the official clinical practice journal of the American Gastroenterological Association*. 2006 Oct;**4**(10):1248–54.
  12. Rutgeerts P, Van Assche G, Vermeire S. Optimizing anti-TNF treatment in inflammatory bowel disease. *Gastroenterology*. 2004 May;**126**(6):1593–610.
  13. Vande Casteele N, Gils A. Pharmacokinetics of anti-TNF monoclonal antibodies in inflammatory bowel disease: Adding value to current practice. *Journal of clinical pharmacology*. 2015 Mar;**55** Suppl 3(May 2014):S39-50.
  14. Perry M, Bewshea C, Brown R, So K, Ahmad T, McDonald T. Infliximab and adalimumab are stable in whole blood clotted samples for seven days at room temperature. *Annals of clinical biochemistry*. 2015 Nov;**52**(Pt 6):672–4.
  15. Perry, M., Bewshea, C., Brown, R., So, K., Ahmad, T., & McDonald, T. (2015). *Infliximab and adalimumab are stable in whole blood clotted samples for seven days at room temperature*. *Annals of Clinical Biochemistry*. Epub ahead of print.

**Used symbols:**

	Temperature limitation		Catalogue number
	For research use only		To be used with
	Manufacturer		Contains sufficient for <n> tests
	Lot number		Use by
	Attention		Consult instructions for use
	Consult specification data sheet		Irritant